Hand Protection Solutions for Critical Environments



Touch N Tuff® 92-605



WWW.ANSELL.EU



Validationpack



Touch N Tuff[®] 92-605 Production Location: Ansell Lanka, Biyagama, Sri Lanka Ansell Thailand, Bangkok, Thailand

Contents:

- 00 Table of contents
- 01 Product Datasheet
- 02 Technical Datasheet
- 03 CE Declaration of Product Conformity
- 04 CE Certificate
- 05 Viral Penetration ISO 16604:2004 or ASTM 1671:95
- 06 Statement of Freedom of Latex
- 07 Permeation Breakthrough times EN 374-3
- 08 Comparison of chemical resistance between several polymers
- 09 Packaging data
- 10 ACPP-Ansell Cytostatic penetration
- 11 Quality management ISO9001:2000
- 12 Declaration EN 455
- 13 Food Declaration of Product Conformity
- 14 GMP Food Declaration
- 15 Health statement
- 16 Instructions for Use
- 17 Ansell's pharmaceutical range for laboratory and cleanroom environments




Ansell Healthcare Europe N.V. Riverside Business Park, Block J - Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 http://www.ansell.eu E-mail info@ansell.eu

ISO 9001 Certificate Number FM 40130

Touch N Tuff® 92-600 92-605

| | | | CHEMICAL & LIQUID PROTECTION | | | | | | |
|---------|-------------------|------------------|------------------------------|----------------|-----------------|----------------------|--|--------------|--|
| STYLE # | LINER MATERIAL | CUFF STYLE | COLOUR | GRIP DESIGN | THICKNESS MM | POW- DER- FREE | EN SIZE | LENGTH MM | PACKAGING |
| 92-600 | Not applicable | Rolled Beaded | Green | Smooth | 0.12 | Yes | 6½-7, 7½-8, 8½-9, 9½-10 | 240 | 100 gloves in a box, 10 boxes in a carton |
| 92-605 | Not applicable | Rolled Beaded | Green | Smooth | 0.12 | Yes | 6½-7, 7½-8, 8½-9, 9½-10, 10½-11 | 300 | 100 gloves in a box, 10 boxes in a carton |





Proven splash resistance against hazardous chemicals

PRIMARY INDUSTRIES





IDEAL APPLICATIONS

- Chemical handling
- Laboratory analysis
- Pharmaceuticals
- Paint and spray shops Printing industry
- Electronics Light assembly tasks
- Light assembly of oil-coated pieces
- Glass manufacturing
- Handling of cytostatics
- Intricate parts handling •




DESCRIPTION

- Further testing of the Touch N Tuff® by a certified body on an even wider range of chemicals confirms that it resists a greater variety of industrial chemicals for longer periods than any other
- Manufactured from nitrile, Touch N Tuff® offers up to four times the puncture resistance of comparable natural-latex gloves, and three times the resistance of similar neoprene gloves.
- It contains no natural rubber proteins, which means no risk of Type I allergies for the wearer. Primary skin irritation studies and Insult Patch tests have also shown no evidence of risk from
- With a unique "Thin Nitrile Technology" formulation, this glove
- offers easy donning and strong grip in wet or dry conditions.
- The glove is highly versatile and suitable for many different uses.
- AQL 1.5 (EN374)

ייק





Touch N Tuff® is anti-staticity tested (EN 1149-1 & 1149-3) and meets the requirements defined in EN 1149-5





Touch N Tuff[®]

Product protection

Orig 06-07-2004 Rev 26-07-2010

Product description

Blue-green unlined nitrile rubber disposable gloves. Thickness: min. 0.10 mm Typical 0.12 mm

Glove material

Acrylonitrile-butadiene rubber

Possible Harmful ingredients

Sensitizers:

- Zinc mercaptobenzothiazole
- Zinc dibutyldithiocarbamate

Only a very small number of users may be sensitised to any of these ingredients and hence may develop irritant and/or allergic reactions.

No Substances of Very High Concern (as per REACH Regulation 1907/2006)

Properties

Excellent physical properties Excellent comfort Excellent puncture resistance Free of silicone and plasticisers No latex proteins: low risk for skin allergies. Very good chemical protection against solvents: please contact Ansell Protective Products.

EN Test Data

Abrasion resistance:Performance level 0Blade cut resistance:Performance level 0 (index 1.1)Tear resistance:Performance level 0 (0.2 N)Puncture resistance:Performance level 0 (16.3 N)Anti-staticity:-Surface resistance: $5.5 \times 10^9 \Omega$
-Vertical resistance: $1.5 \times 10^8 \Omega$ Liquidproofness:Acceptable Quality Level 1.5or EN performance level 2Chemical permeation:

| Туре | Reference | (EN) Sizes | Length |
|-------------|-----------|---|--------|
| | 92-600 | 6 ^{1/2} -7, 7 ^{1/2} -8, 8 ^{1/2} -9, 9 ^{1/2} -10 | 240 mm |
| Powder-free | 92-605 | 6 ^{1/2} -7, 7 ^{1/2} -8, 8 ^{1/2} -9, 9 ^{1/2} -10, 10 ^{1/2} -11 | 300 mm |

Ansell Healthcare Europe N.V. (European Head Office) Riverside Business Park, Block J Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 Ansell (UK) Limited 30-32 Mariner Lichfield Road Industrial Estate Tamworth Staffordshire B797UL, United Kingdom Tel. +44 1827 302100 Fax +44 1827 302101 http://www.ansell.eu E-mail info@ansell.eu



ISO 9001 Certificate Number FM 40130



Ansell Healthcare Europe N.V.

Riverside Business Park

Boulevard International 55

Block J B-1070 Brussels Tel. 32 (0)2-528 74 00 Fax 32 (0)2-528 74 01

EC DECLARATION OF PRODUCT CONFORMITY

Category III

The manufacturer, established in the European Economic Community:

ANSELL HEALTHCARE EUROPE N.V. RIVERSIDE BUSINESS PARK, BLOCK J BOULEVARD INTERNATIONAL 55 B-1070 BRUSSELS

declares that the PPE described hereafter:

Touch N Tuff® 92-605 Image: Construction of the second s

is in conformity with the provisions of the Council Directive 89/686/EEC and with the European harmonised standards EN420: 2003, EN421: 1994, EN374: 2003, and is identical to the PPE which is subject to the EC Type Examination certificate number 03210376 issued by the Notified Body:

CENTEXBEL (0493) TECHNOLOGIEPARK 7 B-9052 ZWIJNAARDE

is subject to the procedure set out in Article 11 point A of Directive 89/686/EEC under the supervision of the Notified Body

CENTEXBEL (0493) TECHNOLOGIEPARK 7 B-9052 ZWIJNAARDE

Stor On

Wednesday, February 02, 2011 Guido VAN DUREN Director Technical Services ANSELL HEALTHCARE EUROPE N.V.



In application of the directive 89/686/EEG of 21 December 1989 concerning the harmonisation of the Member States legislation relative to personal protective equipment, Centexbel Notified body 0493 authorised by decree AV/OA235/ST dated 94-05-25 of the Ministry of Employment and Labour has issued

to: Ansell Healthcare Europe nv Riverside Business Park - Spey House

Internationale Laan 55 1070 Brussels Belgium

CE TYPE EXAMINATION CERTIFICATE

Nr. 03208213

This CE Type examination certificate is valid until 12/07/2015

for: The gloves Touch N Tuff 92-500 and 92-600

The personal protective equipment above mentioned satisfies the applicable essential safety requirements of the Directive.

For the argumentation, the following standards are used:EN 420:2003:Protective gloves - general requirementsEN 374-1:2003 (low):Protective gloves against chemicals and micro-organisms

This is PPE of category III, subject to regular checks in accordance with article 11 of the European PPE directive. In agreement with the manufacturer's choice random checks shall be carried out to assess the quality of the final product (art.11A). The manufacturer must be able, on request, to present the test report of this quality control check. A first quality control check shall be performed at the latest on 31.12.2011 and at least be repeated with intervals of one year.

The technical file is registered with number 2726

Inge De Witte

certification officer

20 May 2008

Re-issued 12 July 2010

10 Jan Laperre

general manager





Annex to certificate 03208213

The gloves Touch N Tuff 92-500 and 92-600 fulfil the requirements of EN 374-1:2003 (low):

to use the following pictogram





CENTEXBEL GENT Technologiepark 7, BE-9052 Zwijnaarde (Gent) tel. + 32 9 220 41 51 fax + 32 9 220 49 55 e-mail gent@centexbel.be BTW BE 0459 218 289 - fin.rek. 210-0472965-45 - IBAN BE 44 2100 4729 6545

Ansell Healthcare Europe

c/o Mr Van Duren

Riverside Business Park, Spey House Boulevard International 55 1070 Bruxelles Belgique

your notice of your ref. 22/04/2008

our ref. YR/cr/08/250 date Chaineux, 22 May 2008

Analysis Report Nr. 8248

Required test :

ISO 16604 - procedure B : 2004

Description of the samples :

| Identification number | Information given by the customer | Date of receipt |
|--------------------------|-----------------------------------|-----------------|
| V800374 | Touch N Tuff 92-600/605 | 24/04/08 |
| V800375 | TNT Blue 92-670/665 | 24/04/08 |

Yvette Rogister Order responsible

For further information please contact our sectorial adviser : Mark Croes.

This report runs to 6 pages and may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel. The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples.



VAT BE 459 218 289 Fin. Acc. 210-0472965-45 CENTEXBEL-VERVIERS Avenue du Parc, 38 B-4650 Hervz (Chaineux) Tel. + 32 87 32 24 30 • Fax + 32 87 34 05 18 e-mail chaineux@centexbel.be

IBAN BE44 2100 4729 6545 CENTEXBEL-BRUSSELS Montoyerstraat 24 B2 BE-1000 Brussels Tel. + 32 2 287 08 30 • Fax + 32 2 230 68 15

+ * *

our reference date page YR/cr/08/250 22 May 2008 2/6

Reference : V800374 V800375

Touch N Tuff 92-600/605 TNT Blue 92-670/665

ISO 16604 : 2004

Clothing for protection against contact with blood and body fluids – Determination of the resistance of protective clothing materials to penetration by blood-borne pathogens – Test method using Phi-X174 bacteriophage

Date of ending the test : 22/05/2008

1. Method

ISO 16604 : 2004 : Clothing for protection against contact with blood and body fluids – Determination of the resistance of protective clothing materials to penetration by bloodborne pathogens – Test method using Phi-X174 bacteriophage

A specimen is subjected to a nutrient broth containing a virus for a specified time and pressure sequence. Visual detection of penetration is supplemented with an assay procedure that will detect viable viruses which penetrate the material even when liquid penetration is not visible. Any evidence of viral penetration for a sample constitutes failure.

In the method, the specimen is inserted in the penetration cell with the normal outside surface of the textile toward the cell reservoir which is further filled with the Phi-X174 bacteriophage challenge suspension.

The other face is, for procedures B and D, in contact with a retaining screen (which assures a good bearing of the textile during the pressure application).

The pressure application procedure depends on the type of material, and the applied pressure is chosen according to the result obtained with ISO 16603 (highest pressure with no visible penetration in ISO 16603 used for ISO 16604).

| our reference | date | page |
|---------------|-------------|------|
| YR/er/08/250 | 22 May 2008 | 3/6 |

2. Results

General information

| ٠ | Sampling : | sampling in 3 gloves, in the palm and the back of hand. |
|---|--------------------------------|---|
| ٠ | Number of test specimens : | 3 |
| • | Test specimen dimension : | 7.5 X 7.5 cm ² |
| • | Conditioning of the sample : | 24 hours at $21 \pm 5^{\circ}$ C and 60 ± 10 % relative humidity. The sample is not tested in conditioned atmosphere but directly after conditioning. |
| ٠ | Sterilisation of the samples : | None |
| • | Side in contact with the | |
| | bacteriophages suspension : | external side |
| ٠ | Test procedure : | PROCEDURE B with the following pressure and time sequence : - 0 kPa for 5 min, |
| | | - followed by 14 kPa for 1 min, |
| | | - followed by 0 kPa for 4 min. |
| | | Use of a metal square mesh screen (open area $>50\%$, limiting deflection of the sample limited to ≤ 5.0 mm). |
| ٠ | Surface tension of the | |
| | bacteriophages suspension: | 42 +/- 2 dynes/cm. |
| ٠ | Used bacteriophage : | Bacteriophage Phi-X174 (ATCC 13706-B1) |
| ٠ | Host bacteria : | Escherichia coli (ATCC 13706) |
| | | |

| our reference | date | page |
|---------------|-------------|------|
| YR/cr/08/250 | 22 May 2008 | 4/6 |

| Centexbel reference of the sample | | | V80 | 0374 | | |
|--|----------------|----------|---------------------|---------------------------------|--------------|----------|
| Pressure | | | 141 | kPa | | |
| Type of sample | | | Green | Gloves | | |
| Sample thickness | | | Not rec | quested | | |
| Sample weight | | | Not rec | quested | | |
| Paraffin-sealed edges | | | N | lo | | |
| Sterilisation type | | | No | one | | |
| Procedure selected | | | I | В | | |
| Inhibition test : ratio | | | 1 | .6 | | |
| Number of plaques per settle plates during the different steps of the test (PFU) | <u>Samp. 1</u> | Samp.2 | Samp.3 | Contr. ⁽¹⁾ conta. | <u>Contr</u> | Contr.+ |
| During the filling of the cell | 0 | 0 | 0 | 0 | 0 | 0 |
| During the rinsing and the | 0 | 0 | 0 | 0 | 0 | 0 |
| draining of the cell | _ | _ | <u> </u> | | • | <u> </u> |
| During the test During the initial titer | 0 | 0 | 0 | 0 | 0 | 0 |
| determination | , v | - | - | - | - | - |
| During the final titer | - | - | - | - | - | 0 |
| determination | | | | | | |
| Titer of the initial bacteriophages suspension | | | Before filling | g the first cell | | |
| Titre (PFU/ml) | | | 2 10 ⁸ 1 | PFU/ml | | |
| Titer of the bacteriophages | | | After the | e last cell | | |
| suspension after the test | | | | | | |
| (drained from the last cell) | | | 2 6 108 | PFU/ml | | |
| Titre (PFU/ml) | | | | | | |
| <u>Results</u> | Sample 1 | Sample 2 | Sample 3 | Contr. ⁽¹⁾ conta. | Contr | Contr. + |
| | | | | Contra. | | |
| Number of plaques 1 | 0 | 0 | 0 | 0 | 0 | high |
| Number of plaques 2 | 0 | 0 | 0 | 0 | 0 | high |
| Number of plaques 3 | 0 | 0 | 0 | 0 | 0 | high |

(1): Control of contamination (only when the sample can't be sterilised)

Performed under accreditation in the microbiological lab under the responsibility of Yvette Rogister.

| our reference | date | page |
|---------------|-------------|------|
| YR/er/08/250 | 22 May 2008 | 5/6 |

| Centexbel reference of the sample | | | V80 | 0375 | | |
|--|----------|----------|---------------------|---------------------------------|--------------|----------|
| Pressure | | | 141 | kPa | | |
| Type of sample | | | Blue (| Gloves | | |
| Sample thickness | | | Not rec | quested | | |
| Sample weight | | | Not rec | quested | | |
| Paraffin-sealed edges | | | N | lo | | |
| Sterilisation type | | | No | one | | |
| Procedure selected | | | H | 3 | | |
| Inhibition test : ratio | | | 1.5 | 88 | | |
| Number of plaques per settle plates during the different steps of the test (PFU) | Samp. 1 | Samp.2 | Samp.3 | Contr. ⁽¹⁾ conta. | <u>Contr</u> | Contr.+ |
| During the filling of the cell | 0 | 0 | 0 | 0 | 0 | 0 |
| During the rinsing and the | 0 | 0 | 0 | 0 | 0 | 0 |
| draining of the cell | _ | <u> </u> | 0 | | 0 | <u> </u> |
| During the test During the initial titer | 0 | 0 | 0 | 0 | 0 | 0 |
| determination | , v | - | - | - | - | - |
| During the final titer determination | - | - | - | - | - | 0 |
| Titer of the initial | | | Before filling | g the first cell | | |
| bacteriophages suspension Titre (PFU/ml) | | | | PFU/ml | | |
| Titer of the bacteriophages | | | After the | e last cell | | |
| suspension after the test | | | | | | |
| (drained from the last cell.) Titre (PFU/ml) | | | 2 3 10 ⁸ | PFU/ml | | |
| Results | Sample 1 | Sample 2 | Sample 3 | Contr. ⁽¹⁾ | Contr | Contr. + |
| Results | sample 1 | Sample 2 | sample 5 | conta. | conu | Conu. + |
| | | | | | | |
| Number of plaques 1 | 0 | 0 | 0 | 0 | 0 | 18 |
| Number of plaques 2 | 0 | 0 | 0 | 0 | 0 | 21 |
| Number of plaques 3 | 0 | 0 | 0 | 0 | 0 | 17 |

Performed under accreditation in the microbiological lab under the responsibility of Yvette Rogister.

| our reference | date | page |
|---------------|-------------|------|
| YR/er/08/250 | 22 May 2008 | 6/6 |

Summary : ISO 16604 - procedure B

| Sample | Replicate | Results |
|---------|-----------|---------|
| V800374 | 1 | Pass |
| 14 kPa | 2 | Pass |
| | 3 | Pass |
| V800375 | 1 | Pass |
| 14 kPa | 2 | Pass |
| | 3 | Pass |

Ansell Occupational Healthcare

c/o Mr. Guido Van Duren

Internationale Laan 55 B-1070 Brussels

your notice of 10/03/2006 25/09/2009 your ref.

our ref. YR/sn/09/610 date Chaineux, 28 September 2009

Analysis Report Nr. 6121/B

Modification of analysis report 6121, made on 4 May 2006

Required test :

ASTM F 1671 : 1997

Description of the samples :

| Identification number | Information given by the customer | Date of receipt |
|--------------------------|--|-----------------|
| V600127 | Sol-Vex 37-675 - green rubber gloves | 15/03/2006 |
| V600128 | Sol-Vex 37-645 - green rubber gloves | 15/03/2006 |
| V600129 | Bi-Colour 87-900 lot 0504045509 - green and yellow rubber gloves | 15/03/2006 |
| V600130 | Extra 87-950 lot 0410072309 - black rubber gloves | 15/03/2006 |
| V600131 | Profil Plus 87-850 lot 0310046909 - yellow rubber gloves | 15/03/2006 |

Yvette Rogister Order responsible

For further information please contact our sectorial adviser : Marc Croes.

This report runs to 18 pages and may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel. The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples.

Fin. Acc. 210-0472965-45



VAT BE 459 218 289 CENTEXBEL-VERVIERS Avenue du Parc, 38 B-4650 Hervz (Chaineux) Tel. + 32 87 32 24 30 • Fax + 32 87 34 05 18 e-mail chaineux@centexbel.be IBAN BE44 2100 4729 6545 CENTEXBEL-BRUSSELS Montoyerstraat 24 82 BE-1000 Brussels Tel. + 32 2 287 08 30 • Fax + 32 2 230 68 15

| Identification number | Information given by the customer | Date of receipt |
|-----------------------|--|-----------------|
| V600132 | Universal Plus 87-660 lot 0507618028 - red rubber gloves | 15/03/2006 |
| V600133 | Touch N Tuff 92-600.605 lot 0408106728 - green rubber gloves | 15/03/2006 |
| V600134 | Conform + 69-150 lot 05070322 EP - beige latex gloves | 15/03/2006 |
| V600135 | Micro-Touch Ultra PF lot 05092423EP - beige latex gloves | 15/03/2006 |
| V600136 | BN45L – black apron | 15/03/2006 |
| V600137 | PVC45G – green apron | 15/03/2006 |
| V600138 | VSB-8-HS - white "arm or leg" protective cover | 15/03/2006 |

Yvette Rogister Order responsible

For further information please contact our sectorial adviser : Marc Croes.

This report runs to 18 pages and may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel. The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples.



VAT BE 459 218 289 CENTEXBEL-VERVIERS Avenue du Parc, 38 B-4650 Hervz (Chaineux) Tel. + 32 87 32 24 30 • Fax + 32 87 34 05 18 e-mail chaineux@centexbel.be Fin. Acc. 210-0472965-45

IBAN BE44 2100 4729 6545 CENTEXBEL-BRUSSELS Montoyerstraat 24 82 BE-1000 Brussels Tel. + 32 2 287 08 30 • Fax + 32 2 230 68 15

our referencedatepageYR/sn/09/61028 September 20093/18

References :

V600127 Sol-Vex 37-675 - green rubber gloves

V600128 Sol-Vex 37-645 - green rubber gloves

V600129 Bi-Colour 87-900 lot 0504045509 - green and yellow rubber gloves

V600130 Extra 87-950 lot 0410072309 - black rubber gloves

V600131 Profil Plus 87-850 lot 0310046909 - yellow rubber gloves

V600132 Universal Plus 87-660 lot 0507618028 – red rubber gloves

V600133 Touch N Tuff 92-600 lot 0408106728 – green latex gloves

V600134 Conform + 69-150 lot 05070322 EP – beige latex gloves

V600135 Micro-Touch Ultra PF lot 05092423EP – beige latex gloves

V600136 BN45L – black apron

V600137 PVC45G – green apron

V600138 VSB-8-HS - white "arm or leg" protective cover

ASTM F 1671 – 97 Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System Specimen Exposure Procedure B

Date of ending the test : 02/05/2006

1. Method

• ASTM F1671-97 : Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System - Specimen Exposure Procedure B

In the method, the textile is put into the cell test. One side of the textile is directly in contact with a bacteriophages suspension (phi-X174) (textile side to be tested), the other side is in contact with a retaining screen (which ensures a good bearing of the textile during the pressure application). After assembly, the cell is placed in the ASTM equipment and the pressure is applied. The bacteriophage penetration through the textile is induced by this pressure application. The application procedure is the following one : 5 minutes of contact without pressure application, 1 minute of contact at 13.8 kPa, 54 minutes of contact without pressure application.

A the end of the test, the sample surface which has not been in contact with the bacteriophages suspension is rinsed. The liquid of rinsing is then put on an agar plate previously inoculated with *Escherichia coli* (host bacteria of the used bacteriophage).

The plates are incubated during 24 hours at 37°C and the presence of plaques (lysis of the bacteria cells) on the agar surface signified that the bacteriophage is passed through the sample.

Performed under accreditation in the microbiological lab under the responsibility of Yvette Rogister

Analysis Report Nr 6121/B

 our reference
 date
 page

 YR/sn/09/610
 28 September 2009
 4/18

The detection of only one plaque constitutes a failure of the textile. Results are expressed in the form : PASS or FAIL Test .

2. <u>Results</u>

Sampling method

- For the gloves, we have cut the samples from the inside and the back of the hand. We have cut in the centre a sample of 7.5 cm x 7.5 cm
- For the other samples, we have cut samples of 7.5 cm x 7.5 cm at random

General information

- Sterilization of the sample : Ethylene oxide
- Pressure at which the sample is submitted : 13.8 KPa 1 minute
- Used bacteriophage : Bacteriophage Phi-X 174 (ATCC 13706-B1)
- Host bacteria : Escherichia coli (ATCC 13706)
- Surface tension of the bacteriophage suspension : 42 +/- 2 dynes/cm.
- Test is repeated three times
- Tested side : outer side

e.

Analysis Report Nr 6121/B

| our reference | date | page |
|---------------|-------------------|-------|
| YR/sn/09/610 | 28 September 2009 | 11/18 |

| Centexbel reference of the sample | | | V60 | 0133 | | | | | |
|--|---------------------|-----------------|-----------------|--|--------------|-----------------|--|--|--|
| Type of sample | Green rubber gloves | | | | | | | | |
| Sample thickness | Not asked | | | | | | | | |
| Sample weight | | | Not | asked | | | | | |
| Paraffin-sealed edges | | | Ν | 10 | | | | | |
| Sterilization type | | | Ethyle | ne oxide | | | | | |
| Inhibition test : ratio | | | | 3 | | | | | |
| Number of plaques per settle plates during the different steps of the test (PFU) | <u>Samp. 1</u> | <u>Samp.2</u> | <u>Samp.3</u> | <u>Contr.⁽¹⁾</u> <u>conta.</u> | <u>Contr</u> | <u>Contr.+</u> | | | |
| During the filling of the cell | 0 | 0 | 0 | , | 0 | 0 | | | |
| During the rinsing and the | 0 | 0 | Ő | 1 | 0 | 0 | | | |
| draining of the cell | | | | , | | | | | |
| During the test During the initial titer | | 0 | 0 | - | 0 | 0 | | | |
| determination | ľ | | _ | _ | | _ | | | |
| During the final titer determination | - | - | - | - | - | 0 | | | |
| <u>Titer of the initial</u> bacteriophages suspension | | | Before filling | g the first cell | <u>l</u> | · | | | |
| Titer (PFU/ml) | | | 1.1 108 | PFU/mi | | | | | |
| Titer of the bacteriophages suspension drained from the | | | After the | e last cell | | | | | |
| last cell after the test Titer (PFU/ml) | | | 1.5 108 | PFU/ml | | | | | |
| <u>Results</u> | <u>Samp. 1</u> | <u>Sampl. 2</u> | <u>Sampl. 3</u> | <u>Contr.⁽¹⁾</u> <u>conta.</u> | <u>Contr</u> | <u>Contr. +</u> | | | |
| Number of plaques 1 | 0 | 0 | 0 | / | 0 | A lot | | | |
| Number of plaques 2 | 0 | 0 | 0 | 1 | 0 | A lot | | | |
| Number of plaques 3 | 0 | 0 | 0 | / | 0 | A lot | | | |
| Retaining screen specifications | Metallic screen | | | | | | | | |

(1): Control of contamination (only when the sample can't be sterilized)

Performed under accreditation in the microbiological lab under the responsibility of Yvette Rogister

Analysis Report Nr 6121/B

| our reference | date | page |
|---------------|-------------------|-------|
| YR/sn/09/610 | 28 September 2009 | 18/18 |

| Sample | Replicate | Results |
|---------|-----------|---------|
| V600133 | 1 | Pass |
| | 2 | Pass |
| | 3 | Pass |
| V600134 | 1 | Pass |
| | 2 | Pass |
| | 3 | Pass |
| V600135 | 1 | Pass |
| | 2 | Pass |
| | 3 | Pass |
| V600136 | 1 | Pass |
| | 2 | Pass |
| | 3 | Pass |
| V600137 | 1 | Pass |
| | 2 | Pass |
| | 3 | Pass |
| V600138 | 1 | Pass |
| | 2 | Pass |
| | 3 | Pass |

Performed under accreditation in the microbiological lab under the responsibility of Yvette Rogister





Touch N Tuff[®] 92 -600.605

Herewith, I confirm that the Ansell Touch N Tuff[®] 92 -600.605 gloves are made from nitrile. They do not contain natural rubber and hence are free from any latex proteins which are known to cause allergens to sensitized people.

TVICIPODEE Fins

Ann Van den Borre Technical Manager Ansell Healthcare Europe



ISO 9001 Certificate Number FM 40130

Ansell Healthcare Europe N.V.

Riverside Business Park, Block J - Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 <u>http://www.ansell.eu</u> E-mail <u>info@ansell.eu</u>

Glove :

Touch N Tuff® 92-605

| Chemical Agent | Breakthrough Time | Protection Index | CAS Number | Notified Body | EN Standard |
|--|----------------------|---------------------|------------|------------------|-------------|
| 1,1,1-trichloro-2-methyl-2- popyl alcohol in Peanut oil | > 480 | 6 | | Centexbel | 374-3:2003 |
| 1,2-dichloroethane | < 20 | 0 | 107-06-2 | Centexbel | 374-3:2003 |
| 1-lododecane | > 60 | 3 | 2050-77-3 | Force Technology | 374-3:2003 |
| 1-Methoxy-2-Propanol | 14 | 1 | 107-98-2 | Centexbel | 374-3:2003 |
| Acetic Acid, Glacial | 7 | 0 | 64-19-7 | Centexbel | 374-3:2003 |
| Acetone | 0.5 | 0 | 67-64-1 | Force Technology | 374-3:2003 |
| Acetonitrile | < 5 | 0 | 75-05-8 | Centexbel | 374-3:2003 |
| Acetonitrile 73% + Methyl Alcohol 25% + Ammonia 2% | 1 | 0 | | Centexbel | 374-3:2003 |
| Acrylamide, 40% | > 480 | 6 | 79-06-1 | Force Technology | 374-3:2003 |
| Acrylic Acid | < 5 | 0 | 79-10-7 | Centexbel | 374-3:2003 |
| Allylchloride | 70 | 3 | 107-05-1 | Centexbel | 374-3:2003 |
| Ammonium Hydroxide, 25% | 29 | 1 | 1336-21-6 | Centexbel | 374-3:2003 |
| Anioxyde 1000 | > 480 | 6 | | Force Technology | 374-3:2003 |
| Benzyl Alcohol | 10 | 0 | 100-51-6 | Centexbel | 374-3:2003 |
| Bromochloromethane | 88 | 3 | 74-97-5 | Centexbel | 374-3:2003 |
| Butyl Alcohol | > 480 | 6 | 71-36-3 | Centexbel | 374-3:2003 |
| Cacodylic acid Sodium salt buffer 0,1M | > 480 | 6 | | Centexbel | |
| Caffeine 1.6% | > 480 | 6 | 58-08-2 | Centexbel | 374-3:2003 |
| Carbon disulfide | < 5 | 0 | 75-15-0 | Centexbel | 374-3:2003 |

| Permeation breaktrough times according to EN374-3:2003 (minutes) | | | | | | | | | |
|--|----------|------------|-----------------|---------|---------|-------|--|--|--|
| 0 | 4 | 5 | 6 | | | | | | |
| < 10 | 10-30 | 30-60 | 60-120 | 120-240 | 240-480 | > 480 | | | |
| Not recommended | Medium p | protection | High protection | | | | | | |

Data given in the table above are based on results of laboratory tests performed on the palm area of the glove or are based on extrapolations from the results of laboratory tests. These tests were run using standard test methods that may not adequately replicate any specific conditions of end use. Because Ansell has no detailed knowledge or control over the conditions of end use, any of these data must be advisory only, and Ansell must decline any liability.

Ansell Healthcare Europe N.V.





Permeation Breakthrough Times

| Material & Thickness | Polychloroprene (0,12mm) | Nitrile (0,12 mm) | Latex (0,12 mm) |
|-------------------------------|-----------------------------|-------------------|-----------------|
| | | | |
| Productname | NeoTouch® | Touch N Tuff® | Conform®+ |
| | Permeation breakthrough tim | ne (minutes) | |
| | | | |
| 70% Isopropanol (CAS 67-63-0) | >60 | >60 | <10 |
| 70% Methanol (CAS 67-56-1) | <10 | <10 | <10 |
| 70% Ethanol (CAS 64-17-5) | 14 | 27 | <10 |
| Disinfectants based on | | | |
| quaternary ammonium | | | |
| compounds | >480 | >480 | >480 |
| 70% Ethanol with 0,125% H2O2 | >10 | >10 | <10 |

| | Permeatio | on breakthroug | h times accordir | ig to EN374 (minut | es) | |
|------|-----------|----------------|------------------|--------------------|---------|-------|
| 0 | 1 | 2 | 3 | 4 | 5 | 6 |
| < 10 | 10-30 | 30-60 | 60-120 | 120-240 | 240-480 | > 480 |

Medium protection

High protection

Splash protection


Not recommended

Data given in this recommendation are based on results of laboratory tests performed on the palm area of the glove or are based on extrapolations from the results of laboratory tests. These tests were run using standard test methods that may not adequately replicate any specific conditions of end use. Because Ansell has no detailed knowledge or control over the conditions of end use, any of these data must be advisory only, and Ansell must decline any liability.

| Style | Size | Product Name | ASPN | Tariff Code | UOM | Alt. UOM | Quantity | | | Case height (cm) | Case weight (kg) | Bar code for the smallest packaging unit | Bar code for the inner polybag or the shipper |
|--------|---------|-------------------------|--------|-------------|------|----------|----------|----|----|------------------------|------------------------|--|--|
| 92-605 | 6.5-7 | Touch N Tuff® Long Cuff | 588954 | 1015190000 | Case | Piece | 1000 | 47 | 26 | 30 | | 0076490483463 | 20076490483467 |
| 92-605 | 7.5-8 | Touch N Tuff® Long Cuff | 588955 | 4015190000 | Case | Piece | 1000 | 47 | 26 | 30 | | 0076490483456 | 20076490483450 |
| 92-605 | 8.5-9 | Touch N Tuff® Long Cuff | 588956 | 4015190000 | Case | Piece | 1000 | 47 | 26 | 30 | | 0076490483449 | 20076490483443 |
| 92-605 | 9.5-10 | Touch N Tuff® Long Cuff | 588957 | 4015190000 | Case | Piece | 1000 | 47 | 26 | 30 | | 0076490483432 | 20076490483436 |
| 92-605 | 10.5-11 | Touch N Tuff® Long Cuff | 588958 | 1015190000 | Case | Piece | 1000 | 47 | 26 | 30 | | 0076490482671 | 20076490482675 |

ACPP: <u>A</u>nsell <u>C</u>hemical <u>P</u>ermeation <u>P</u>rogram

Ansell invests for your protection and information





In partnership with the Université Catholique de Louvain Brussels - Belgium



Why the ACPP?

According to the World Health Organization, between now and 2020 the cancer rate is going to increase by 50 % to about 15 million new cases a year worldwide.

This predicted sharp increase in new cases will mainly be due to the existence of steadily ageing populations in both developed and developing countries, but also to the current prevalence of smoking and the rise in numbers of those adopting unhealthy lifestyles. One of the most frequently used treatments for cancer is chemotherapy. Unfortunately it is well known that cytostatic agents are potentially hazardous for the manipulator (pharmacist, assistant, nurse, surgeon (HIPEC technique) etc.).

Cytostatics are known to be:

- Mutagenic
- Carcinogenic
- Teratogenic.


Some of the other potential effects are:

- Decrease in fertility
- Foetal malformation
- Abortion
- Extra uterine pregnancy
- High toxicity for certain organs.

ACPP project goals:

- To contribute to better protection for healthcare personnel handling cytostatics
- To present results corresponding to real life conditions or, better still, to worst case scenarios (dynamic treatment & cytostatic concentration)
- To be innovative & bring very practical answers to our customers

Permeation: a very complex process




Permeation depends on many parameters and therefore the results obtained cannot be extended to any other material, brand or competitive product.

Dynamic Permeation Device

In order to ensure your maximum protection when handling cytostatics, Ansell has invested in a totally new permeation assessment protocol.

Because gloves are made to be used under dynamic physical conditions (stretching, tension, rubbing etc.) we thought at Ansell that it was our duty to test some of our products for cytostatic permeation under the same constraints.

In partnership with the Université Catholique de Louvain, Brussels, Belgium, we have developed a UNIQUE dynamic permeation device. After dynamic permeation simulation, aliquots of the collecting medium are analysed by LC-MS/MS* or HPLC-DAD*





Dynamic permeation device

The molecules tested are the following

| | Cytostatic | Brand name | Company | Analytical method | Concentra- tion | LOD (Detection limit) | LOQ (Quantification limit) | Mol. Weight | Log P |
|----|-------------------------|----------------|----------------------|-------------------|--------------------|-----------------------------|----------------------------------|----------------|--------|
| 1 | Carmustine (25°C) | Nutrimon BICNU | Almirall-Prodesfarma | HPLC-DAD | 3.0 mg/ml | 59.02 ng/ml | 196.73 ng/ml | 214.0 | 1.50 |
| 2 | Cisplatin (25°C) | Platinol | Bristol-Myers Squibb | HPLC-DAD | 1.0 mg/ml | 49.70 ng/ml | 165.67 ng/ml | 300.1 | 300.1 |
| З | Cyclophosphamide (25°C) | Endoxan | AstaMedica | LCMS/MS | 20.0 mg/ml | 1.95 ng/ml | 31.60 ng/ml | 261.1 | 0.60 |
| 4 | Cytarabine (25°C) | Cytosar | Pharmacia & UpJohn | LCMS/MS | 100.0 mg/ml | 0.97 ng/ml | 4.93 ng/ml | 243.2 | - 2.50 |
| 5 | Docetaxel (25°C) | Taxotere | Aventis Pharma | LCMS/MS | 10.0 mg/ml | 3.11 ng/ml | 17.32 ng/ml | 807.8 | NA |
| 6 | Doxorubicin (25°C) | Adriblastina | Pharmacia | LCMS/MS | 2.0 mg/ml | 35.26 ng/ml | 55.10 ng/ml | 543.5 | 1.30 |
| 7 | Etoposide (25°C) | Vepesid | Bristol-Myers Squibb | HPLC-DAD | 20.0 mg/ml | 63.06 ng/ml | 210.19 ng/ml | 588.5 | 0.60 |
| 8 | 5-Fluorouracil (25°C) | Fluoroblastine | Pharmacia & UpJohn | HPLC-DAD | 50.0 mg/ml | 37.47 ng/ml | 124.92 ng/ml | 130.1 | - 1.00 |
| 9 | lfosfamide (25°C) | Holoxan | AstaMedica | LCMS/MS | 100.0 mg/ml | 4.11 ng/ml | 45.11 ng/ml | 261.1 | NA |
| 10 | lrinotecan (25°C) | Campto | Aventis Pharma | LCMS/MS | 20.0 mg/ml | 16.32 ng/ml | 30.73 ng/ml | 586.6 | NA |
| 11 | Methotrexate (25°C) | Ledertrexate | Wyeth Lederle | LCMS/MS | 25.0 mg/ml | 1.41 ng/ml | 12.85 ng/ml | 454.4 | - 1.80 |
| 12 | Thiotepa (25°C) | Ledertepa | AstaMedica | LCMS/MS | 10.0 mg/ml | 3.06 ng/ml | 17.36 ng/ml | 189.2 | 0.50 |
| 13 | Vinorelbine (25°C) | Navelbine | Pierre Fabre | LCMS/MS | 10.0 mg/ml | 6.48 ng/ml | 43.72 ng/ml | 779.9 | 16.00 |
| 14 | Mitomycin C (43°C) | Mitomycin C | Nycomed Belgium | HPLC-DAD | 0.4 mg/ml | 12.80 ng/ml | 42.50 ng/ml | 318.0 | NA |
| 15 | Oxaliplatin (43°C) | Eloxatin | Sanofi-Synthelabo | HPLC-DAD | 5.0 mg/ml | 19.00 ng/ml | 64.00 ng/ml | 397.3 | NA |
| 16 | Paclitaxel (37°C) * | Taxol | Bristol-Myers Squibb | HPLC-DAD | 6.0 mg/ml | 37.50 ng/ml | 125.00 ng/ml | 853.9 | NA |
| 17 | Ganciclovir (37°C) | Cymevene | Roche | LCMS/MS | 50.0 mg/ml | 6.00 ng/ml | 20.00 ng/ml | 255.2 | - 1.60 |

* LC-MS/MS: Liquid Chromatography Tandem Mass Spectrometry

* HPLC -DAD: High Pressure Liquid Chromatography - Diode Array Detection

| | | Gloves | | CE G | loves | | L | ab Glove | S |
|---|----------|----------------------|------------------------------------|--|--|---|-----------------------|----------------------------|----------------------|
| - | | atics Ition chart | DermaShield® 73-701 (Neopr.) | AccuTech® Sterile UltraClean 91-210 | AccuTech® Sterile Coated 91-250 | AccuTech® Gammex® 91-225 (NRL) | NeoTouch® (Neopr.) | Touch N Tuff® (Nitrile) | Nitrilite® 93-311 |
| | ĺ. | Carmustine | 4 | 4 | 3 | 6 | 6 | 2 | 3 |
| | | Cisplatine | 6 | 6 | 6 | 6 | 6 | 3 | 6 |
| | | Cyclophosphamide | 6 | 6 | 4 | 4 | 5 | 4 | 6 |
| | | Cvtarabine | 6 | 5 | 6 | 6 | 6 | 6 | 6 |
| | | Docetaxel | 6 | 4 | 6 | 6 | 5 | 6 | 6 |
| | | Doxorubicin | 6 | 4 | 6 | 6 | 6 | 6 | 5 |
| | | Etoposide | 6 | 6 | 3 | 2 | 6 | 6 | 6 |
| | min | 5-Fluorouracil | 6 | 5 | 3 | 2 | 6 | 2 | 6 |
| | | | | 4 | - | | | 0 | |
| | | lfosfamide | 6 | | 4 | 4 | 4 | 4 | 6 |
| | ß | Irinotecan | 4 | 4 | 4 | 5 | 4 | 4 | 5 |
| | | Methotrexate | 6 | 6 | 6 | 6 | 6 | 4 | 6 |
| | | Mitomycin C | 6 | 5 | 4 | \geq | \geq | \geq | 6 |
| | | Oxaliplatin | 6 | 6 | 6 | \geq | \geq | | 6 |
| | | Paclitaxel | 6 | 6 | 6 | \geq | \geq | \geq | 6 |
| | | Thiotepa | 6 | 5 | 6 | 5 | 4 | 5 | 6 |
| | | Vinorelbine | 4 | 6 | 6 | 6 | 4 | 6 | 4 |
| | | Ganciclovir | 6 | 6 | 6 | \geq | \geq | \geq | 6 |
| | 1 | | | | | ~ ~ | | | |
| | | Carmustine | 3 | 4 | 2 | 2 | 6 | 2 | 3 |
| | | Cisplatine | 6 | 3 | 3 | 6 | 6 | 3 | 6 |
| | | Cyclophosphamide | 5 | 6 | 4 | 4 | 5 | 4 | 5 |
| | | Cytarabine | 4 | 6 | 6 | 6 | 6 | 6 | 6 |
| | | Docetaxel | 5 | 4 | 5 | 5 | 5 | 5 | 5 |
| | min | Doxorubicin | 6 | 4 | 5 | 6 | 6 | 4 | 5 |
| | | Etoposide | 6 | 4 | 3 | 2 | 6 | 2 | 4 |
| Contact time | 3 | 5-Fluorouracil | 6 | 5 | 3 | 2 | 6 | 3 | 4 |
| | | lfosfamide | 4 | 4 | 4 | 4 | 4 | 4 | 5 |
| ល | | Irinotecan | 4 | 4 | 4 | 4 | 4 | 4 | 5 |
| CL. | 30 | Methotrexate | 5 | 5 | 5 | 6 | 5 | 4 | 6 |
| | | | 4 | 5 | 4 | | 5 | 4 | 5 |
| | 1 | Mitomycin C | | | | > | | > | |
| | | Oxaliplatin | 6 | 5 | 5 | \langle | | \langle | 5 |
| | | Paclitaxel | 4 | 4 | 4 | | | | 4 |
| | | Thiotepa | 5 | 4 | 4 | 5 | 4 | 5 | 6 |
| | | Vinorelbine | 4 | 6 | 6 | 6 | 4 | 6 | 4 |
| | | Ganciclovir | 5 | 5 | 5 | > | \geq | \geq | 5 |
| | 1 | Carmustine | 0 | 0 | 4 | 0 | 6 | 0 | 3 |
| | | | 0 | 2 | | | | <u> </u> | |
| | | Cisplatine | 6 | 3 | 3 | 6 | 6 | 3 | 6 |
| | | Cyclophosphamide | 5 | 5 | 4 | 3 | 4 | 4 | 3 |
| | | Cytarabine | 4 | 6 | 6 | 6 | 6 | 6 | 5 |
| | | Docetaxel | 5 | 4 | 5 | 5 | 5 | 4 | 5 |
| | _ | Doxorubicin | 6 | 4 | 5 | 6 | 6 | 4 | 5 |
| | | Etoposide | 4 | 4 | 3 | | 6 | 2 | 4 |
| | | 5-Fluorouracil | 6 | 5 | 3 | 2 | 4 | 3 | 4 |
| | | lfosfamide | 4 | 4 | 4 | 4 | 3 | 4 | 5 |
| | D | Irinotecan | 4 | 4 | 4 | 4 | 4 | 3 | 4 |
| | Ú | Methotrexate | 5 | 5 | 5 | 6 | 5 | 4 | 6 |
| | | Mitomycin C | 4 | 5 | 4 | \geq | \geq | \geq | 5 |
| | | Oxaliplatin | 4 | 5 | 5 | \sim | \leq | \sim | 5 |
| | | Paclitaxel | 4 | 4 | 4 | \leq | $\overline{}$ | \sim | 4 |
| | | Thiotepa | 5 | 4 | 4 | 5 | 4 | 5 | 5 |
| | | Vinorelbine | 4 | 4 | 5 | 6 | 4 | 5 | 4 |
| | | Ganciclovir | 4 | 5 | 5 | | | | 5 |
| | | GanGiGiUVII | 4 | <u> </u> | <u> </u> | | | | |
| | | Key | to Protection I | evel | | | | perature indicato | r: |
| | | | | | ent | | 25°C 37°C 43 | 3°C | |
| Level 1 : Permeation detected above to 10 times below EN374-3 limit requirement | | | | | | | | | |

| Level 1 : Permeation detected above to 10 times below EN374-3 limit requirement |
|---|
| Level 2 : Permeation detected 10 to 100 times below EN374-3 limit requirement |
| Level 3 : Permeation detected 100 to 1.000 times below EN374-3 limit requirement |
| Level 4 : Permeation detected 1.000 to 10.000 times below EN374-3 limit requirement |
| Level 5 : Permeation detected 10.000 to 100.000 times below EN374-3 limit requirement |
| Level 6 : > Permeation detected at least 100 000 times below EN374-3 limit requirement or |

Results obtained at room temperature. These results are ONLY and SOLELY valid or the glove tested. Any extrapolation to other materials, brands or competitive products would be erroneous and NOT supported by Ansell.

no permeation measured European norm is EN374-3

Ansell Limited is a global leader in barrier protective products. With operations in the Americas, Europe and Asia, Ansell employs more than 11,000 people worldwide and holds leading positions in the natural latex and synthetic polymer glove and condom markets. Ansell Ansell operates in three main business segments: Occupational Healthcare, supplying hand protection to the industrial market; Professional Healthcare, supplying surgical and examination gloves to healthcare professionals; and Consumer Healthcare, supplying condoms and consumer hand protection.

Ansell UK Ltd

Regus Building, Blythe Valley Business Park Solihull West Midlands B90 8AG, United-Kingdom Tel. +44 (o) 1564 711 802 - Fax +44 (o) 1564 711 344 http://www.ansell.eu - E-mail: infouk@eu.ansell.com

Ansell Healthcare Europe N.V. Riverside Business Park, Block J, Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (o) 2 528 74 00 - Fax +32 (o) 2 528 74 01 Fax Customer Service +32 (o) 2 528 74 03 http://www.ansell.eu - E-mail: info@ansell.eu





QUALITY MANAGEMENT SYSTEM - ISO 9001:2008

This is to certify that:

Ansell Lanka (Pvt) Limited **Biyagama Export Processing Zone** Biyagama Sri Lanka

Holds Certificate No: FM 11521

and operates a Quality Management System which complies with the requirements of ISO 9001:2008 for the following scope:

The design and manufacture of sterile and non sterile natural latex and synthetic surgeons gloves.

The provision of a sterilisation service. The design and manufacture of Industrial gloves.

For and on behalf of BSI:

Managing Director, Pharithacha Petchrungruengphol

Originally registered: 01/02/1991

Latest issue: 17/09/2010

Expiry Date: 27/09/2013



Page: 1 of 2

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated <u>online.</u> Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +66(2) 2944889-92. Further clarifications regarding the scope of this certificate and the applicability of ISO 9001:2008 requirements may be obtained by consulting the organization. This certificate is valid only if provided original copies are in complete set.

Thailand Headquarters: 127/25 Panjathani Tower, 20th Floor, Nonsee Road, Chongnonsee, Yannawa, Bangkok 10120. BSI Thailand is a subsidiary of British Standards Institution.

Certificate No: FM 11521

Location

Ansell Lanka (Pvt) Limited Biyagama Export Processing Zone Biyagama Sri Lanka

Registered Activities

The design and manufacture of sterile and non sterile natural latex and synthetic surgeons gloves. The provision of a sterilisation service. The design and manufacture of Industrial gloves.



Latest Issue: 17/09/2010

Expiry Date: 27/09/2013

Page: 2 of 2

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract. An electronic certificate can be authenticated <u>online</u>. Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +66(2) 2944889-92. Further clarifications regarding the scope of this certificate and the applicability of ISO 9001:2008 requirements may be obtained by consulting the organization. This certificate is valid only if provided original copies are in complete set.

Thailand Headquarters: 127/25 Panjathani Tower, 20th Floor, Nonsee Road, Chongnonsee, Yannawa, Bangkok 10120. BSI Thailand is a subsidiary of British Standards Institution.





Touch N Tuff® EN455

Herewith, I confirm that the Touch N Tuff[®] 92-605 gloves are in conformity with the requirements of the European harmonised standard EN455-1:2000 part 1.

Weilbatter han

Ann van den Borre Technical Manager Ansell Healthcare Europe

Ansell Healthcare Europe N.V.

Riverside Business Park, Spey House Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 http://www.anselleurope.com E-mail info@eu.ansell.com



ISO 9001 Certificate Number FM 40130



Ansell Healthcare Europe N.V.


Riverside Business Park Boulevard International 55 Block J B-1070 Brussels Tel. 32 (0)2-528 74 00 Fax 32 (0)2-528 74 01

EC DECLARATION OF PRODUCT CONFORMITY

The manufacturer, established in the European Economic Community:

ANSELL HEALTHCARE EUROPE N.V. RIVERSIDE BUSINESS PARK, BLOCK J BOULEVARD INTERNATIONAL 55 B-1070 BRUSSELS

declares that the PPE described hereafter:



is in conformity with the provisions of the EC-Regulation 1935/2004 and with the EC-Regulation 2023/2006 related to Good Manufacturing Practices (GMP) for Materials and Articles intended to come in contact with Foodstuffs.

Sto O

Guido VAN DUREN Director Technical Services ANSELL HEALTHCARE EUROPE N.V.





Declaration of Conformity

Good Manufacturing Practices for Ansell's materials and articles intended to come in contact with food

Herewith, the undersigned declares that all Ansell gloves, that are intended for contact with Food products, are manufactured in accordance to the following requirements :

1. Regulation 1935/2004

- Gloves are sufficiently inert to preclude substances from being transferred to food in quantities large enough to endanger human health or to bring about an unacceptable change in the composition of the food or a deterioration in its organoleptic properties.

- Gloves are made with only legally acceptable Food-contact ingredients and do not exceed any legal migration levels based on the intended use of the product. Raw materials used in the production of the gloves are specified safe for food contact, have GRAS (Generally Recognized As Safe) status, and are procured from an approved supplier.

2. Regulation 2023/2006:

- Gloves are made as per 'Good manufacturing practice (GMP)' meaning they are consistently produced and controlled to ensure conformity with the applicable rules and applicable quality standards.

This applies to all activities ; from procurement through approved suppliers of materials and all aspects of manufacturing, processing, handling, storage, transport and distribution of the finished article.

- The manufacturing plant has a documented and effective quality assurance system with the purpose of ensuring that materials and articles are of the quality required to ensure conformity with the rules applicable to them and the quality standards necessary for their intended use.

- The qualifications and training of personnel at manufacturing is documented. As well, the manufacturing facility and equipment is designed, cleaned, and maintained as necessary to ensure that in process materials and finished glove products comply with their specifications. Inherent in these requirements are personnel hygiene, pest control, contamination control, prevention of material damage from the environment, etc., etc.

Ansell Healthcare Europe N.V.

Riverside Business Park, Spey House Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 http://www.anselleurope.com E-mail info@eu.ansell.com



ISO 9001 Certificate Number FM 40130



- The manufacturing plant has conducted a formal risk analysis according to an established procedure (ISO 22000 employs HAACP Hazard Analysis Critical Control Points), and evaluate each proposed change to the documented system for its impact on risk to the user of the finished article.

- The manufacturing plant has an effective quality control system and a documented system of tests, inspections, document reviews and formal dispositions on raw materials, in process materials and finished articles. This system include clear decision criteria on materials and articles not meeting specifications.

- The manufacturing's quality control system monitors compliance to Good Manufacturing Practices and correct any failure to comply with GMP without delay. These corrective measures will be made available to competent authorities when relevant.

Ansell shall ensure adherence to the effective implementation of GMP through review of the supplier's internal audit system as described in the ISO 9001 Quality Management System.

- The manufacturing site maintains electronic or paper documents of specifications, manufacturing formulae, and processing necessary to achieve regulatory compliance and product safety.

- Finished articles must be labeled with a unique control number, which relates to specific records held by the manufacturer.

- When requested, the above documents are made available to both Ansell auditors and competent external authorities.

- Restrictions on printing inks, where relevant, are being considered in product formulation and design, and must be applied from the time of printing until final article distribution.

J.C.

Guido VAN DUREN *Technical Services Director* ANSELL HEALTHCARE EUROPE



Ansell Healthcare Europe N.V.

Riverside Business Park, Spey House Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 http://www.anselleurope.com E-mail info@eu.ansell.com





Touch N Tuff[®] 92-600.605

Herewith, I confirm that the Ansell Touch N Tuff[®] 92-600.605 gloves are in compliance with the essential safety & health requirements as outlined by the European Directive 89/686/EEC and relevant EN standards and, as such, can be considered as harmless when the gloves are used according to their intended purpose (= mechanical protection).

This also means that these gloves are compliant to the requirements outlined in EN 420:2003; i.e.

- 1. pH = close to neutral
- 2. The ingredients and materials inside the gloves do not adversely affect the users health.

In addition, we can also confirm that these above mentioned gloves are:

- Free from Chromium VI and Nickel
- Free from any ingredients which are considered as Substances of High Concern as per the REACH Regulation 1907/2006
- Free from any releasable Carcinogenic, Mutagenic and Reprotoxic substances
- Free from any heavy metals
- Free from any azo-or carcinogenic colorants
- Free from any DMF or other releasable solvents

Please consult the technical data sheet of this glove for further details.

VCTPOLLEF FM

Ann Van den Borre Technical Manager Ansell Healthcare Europe



Ansell Healthcare Europe N.V.

Riverside Business Park, Spey House Boulevard International 55 B-1070 Brussels, Belgium Tel. +32 (0) 2 528 74 00 Fax +32 (0) 2 528 74 01 Fax Customer Service +32 (0) 2 528 74 03 http://www.anselleurope.com E-mail info@eu.ansell.com

ISO 9001 Certificate Number FM 40130

| | EN | ES | п | NL | DA | NO | AR |
|----|----|----|----|----|----|----|----|
| | FR | DE | РТ | EL | FL | sv | RU |
| CS | HU | LV | PL | BG | SL | ко | ΤН |
| ET | LT | МТ | RO | SK | TR | MS | ZH |

Europe i

Ansell Healthcare Europe NV Blvd International, 55 1070 Brussels, Belgium www.anselleurope.com

United States

Ansell Healthcare 200 Schulz Drive, Red Bank NJ 07701 ☎ +1 800 800 0444 = +1 800 800 0445 www.ansellpro.com

Japan

Ansell Japan Ltd. Tokyo, Japan **a** +81 3 5805 3781

Malavsia

Ansell Services (Asia) Sdn Bhd **a** +60 3 5541 9797

Canada

105 Lauder Street, Cowansville J2K 2K8 **a** +1 800 363 8340

Australia

Ansell Limited Level 3, 678 Victoria Street, Richmond, Vic, 3121 ☎ +61 1800 337 041 ≞ +61 1800 803 578 www.ansellasiapacific.com



09/2009

INSTRUCTIONS FOR USE CATEGORY III / COMPLEX DESIGN

ANSELL DISPOSABLE AND CRITICAL **ENVIRONMENT GLOVES**

CE&DI

A. Use This Instruction for Use note is to be used in combination with the specific information that is mentioned on or inside Insinstruction for Use note is to be used in combination with the specific information that is mentioned on or inside each packaging enclosure. These gloves are designed as single use gloves and should be disposed of after use. The gloves are liquid proof, and can therefore be used for splash protection against cortain chemicals. They are in conformity with, and are marked per the requirements of the European Directive 89/686/EEC and its amendments. They also comply with the applicable European Standards. Gloves which are accompanied with the pictogram which designates contact with foodstuffs, are also in conformity with the European Regulations 1935/2004 and 2023/2006 as well as with all applicable National Regulations for Food-contact materials

materials. Please ensure the gloves are used only for the designated purpos



Warning! The information given in the pictograms or data provided on chemical resistant breakthrough times is based on lab tests and is therefore advisory only since it does not necessarily reflect the actual duration in the workplace. E0-Type examination cartificate from Cantexbel Belgium (J.D. 0493), Technologiepark 7, B-9052 Zwijipaarde For more detailed information on the giove's performance and to obtain a copy of the Conformity Declaration, please consult Ansell.

B. Precautions for use

B. Precautions for use 1. Before usage, inspect the gloves for any defects or imperfections. If the gloves are ripped or punctured during use, dispose of them immediately. If in doubt, do not use the gloves, get a new pair.
2. If the gloves are used against chemicals, it is essential to keep all chemicals from the skin, even if they are thought to be harmless. Therefore use gloves which are rated with a protection index of 6 or with an excellent degradation resistance rating. In all other cases, the gloves should be used for splash protection or shour contact only. Ensure that chemicals cannot enter via the cuff. For more details regarding chemical suitability, please contact the Ansell Tachylical department. Technical department.

Technical department. 3. Avoid wearing gloves which are dirty on the inside – they may irritate the skin causing dermatitis or worse. 4. Contaminated gloves should be cleaned or washed before removal. 5. The gloves should not be used in applications requiring thermal protection. 6. Disposable glove type versions with a glove length below 260 mm are 'Fit for Special Purpose gloves' because they are to be used to protect the hand ONLY from chemical splashes when handling chemicals. Do not use the gloves when protection in the cuff area is needed. 7. Gloves shall not be used for protection against ionising radiation nor for use in containment enclosures. 8. Not all gloves that are suitable for contact with foodstuffs. To know which restrictions apply and for which specific food Conformity declaration. 9. If gloves are marked, the printed surfaces shall not come in contact with food.

C. Ingredients / Hazardous ingredients Some gloves might contain ingredients which are known to be a possible cause of allergies in sensitised persons, who may develop irritant and/or allergic contact reactions. If allergic reactions should occur, obtain medical advice immediately. For more information, please contact Ansell's Technical department.

D. Care instructions

D. Care instructions
 Storage: Keep away from direct sunlight; store in a cool dry place.
 Keep away from ozone sources or naked flame.
 Cleaning: The gloves are not designed to be laundered.

E. Disposal c. Disposal Used gloves may be contaminated with infectious or other hazardous materials. Dispose of according to Local Authority Regulations. Landfill or incinerate under controlled conditions.

Ansell

Ansell's **pharmaceutical** range for **laboratory** and **cleanroom** environments

For laboratory environments requiring safety and consistent product quality

For applications such as non-sterile drug manufacturing, medical device manufacturing that requires the use of a personal protective glove, biotechnology, laboratory analysis and laboratory work we offer:

| Latex | 69-150 Conform®+ powder-free | | |
|----------|---|--|--|
| Nitrile | 92-600 & 92-605 Touch N Tuff® powder-free | | |
| | 92-670 & 92-665 TNT® Blue | | |
| Neoprene | 25-101 NeoTouch® | | |
| | 25-201 NeoTouch [®] Long | | |

These gloves are ambidextrous, and packed in lots of roopcs in cardboard dispensers. We maintain relevant product infor-

mation on these gloves, such as:

- Technical information
- Lot traceability

For controlled and critical environments that require a low tolerance level of airborne particles

For pharmaceutical and medical device manufacturing, and drug manufacturing applications that do not require sterile conditions, and that require the use of a personal protective glove, we offer:

| Latex | Latex 91-104 AccuTech® Ambi UltraGrip (2) | | |
|----------|---|--|--|
| | 91-325 AccuTech® Gammex® | | |
| | 91-250 AccuTech® Sterile Coated | | |
| Nitrile | 93-311 Nitrilite [®] (2) | | |
| | 93-401 Nitrilite [®] Silky Ultra-Clean (²) | | |
| Neoprene | 73-701, 73-711 & 73-721 DermaShield® | | |

We make technical data and certification information available online, via the Internet, on:

- Lot numbers
- Certificates of analysis
- Historical trend data
- (²) Particle count and extractables Packaging
- No paper contamination
- PE packaging
- IPA-resistant

For sterile environments ISO 5 (Class 100) required for aseptic drug manufacturing

For these applications we offer gloves that are both sterile and clean:

| Latex 91-225 AccuTech® Gammex® Sterile | |
|--|---|
| | 91-250 AccuTech [®] Sterile Coated |
| Neoprene | 73-701 & 73-711 DermaShield® |

In compliance with the relevant ASTM* standards, we can provide you with the following, on request:

- Gamma Irradiation certificatesCertificates of analysis
- (²) Particle count and extractables
- * ASTM is the American equivalent of EN standards, which we refer to when we do not have an EN equivalent.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

For more details about Ansell's pharmaceutical industry product range, please contact:

