

PJD SAFETY SUPPLIES

124 Moore Street, Aughnacloy, Co. Tyrone, BT69 6AA, Northern Ireland, United Kingdom

The Manufacturer or his nominated representative established in the community PJD Safety Supplies

declares that the P.P.E described hereafter

Non-sterile examination and protective Nitrile gloves for single use SIZES; S-XL

We hereby declare under sole responsibility that the CE marked product described above conforms to the essential requirements (Annex I) of the directive for medical devices 93/42/EEC.

The conformity assessment is based on Annex VII. Classification according to rule 5, Annex IX. Applied standards: EN 455-1:2000, EN 455-2:2009+A2:2013 & EN 455-2:2015, EN 455-3:2006 & EN 455-3:2015, EN 455-4:2009, ISO 10993-1:2013, EN ISO 15223-1:2017, EN 1041:2013, EN ISO 14971:2013, DIN ISO 2859-1:2014, , EN ISO 13485:2016

We hereby certify that the CE 0321 marked product described above conforms with the applicable provisions of the directive for personal protective equipment 89/686/EEC and is identical to the personal protective equipment which is the subject of the EC certificate of conformity No. 5213 issued by

SATRA Technology Centre, Wyndham way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom, ID No. 0321

and is subject to the procedure set out in article 11A of Directive 89/686/EEC under the supervision of

SATRA Technology Centre, Wyndham way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom, ID No. 0321

Applied standards: EN 420:2003+A1:2009, EN 374-1:2003, EN 374-2:2003, EN 374-3:2003, EN 388:2003

P Coote

DIRECTOR

Date: 21/04/2020 Issue No: 1